

# CERTIFICATE OF CALIBRATION

## STANDARD CALIBRATION

### BLACK DIAMOND POLICE DEPARTMENT

*This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2005 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2005, that is "As Found" data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.*

<b>DESCRIPTION:</b>	<b>SPEED MEASURING DEVICE</b>		
<b>I.D.:</b>	RP01593	<b>SERIAL NUMBER:</b>	RP01593
<b>MANUFACTURER:</b>	KUSTOM	<b>MODEL NUMBER:</b>	RP-1
<b>DEPARTMENT:</b>		<b>LOCATION:</b>	
<b>ENVIRONMENT:</b>	71.0 °F/51.0 %RH	<b>BASIC ACCURACY:</b>	REFERENCE MFR SPECIFICATIONS
<b>CAL INTERVAL:</b>	12 MONTHS	<b>DUE DATE:</b>	April 07, 2010

#### EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

#### EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were IN TOLERANCE, as defined by the basic accuracy stated above.

#### TUNNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNNING FORK ONE	45296	35 M.P.H.	2540 Hz
TUNNING FORK TWO	43260	65 M.P.H.	4722 Hz

#### STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	12/11/2009
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	12/11/2009

#### PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
SMD101	DOPPLER RADAR / LIDAR CALIBRATION PROCEDURE	NEW	02/01/2006

#### CERTIFICATION NOTES

PERFORMED BY

*Charlie M. Brown*  
 CHARLIE BROWN

CALIBRATION DATE: Tuesday, April 07, 2009